



EU Declaration of Conformity

Name and address of the manufacturer:

Dr. Arabin GmbH & Co. KG, Alfred-Herrhausen-Str. 44, 58455 Witten, Germany

We declare under our sole responsibility that the medical device

Product name	Article Code / REF Nr.	Class
Ring Pessary	RP050 - 100	Ila
Bowl Pessary	SP055 - 095	Ila
Sieve Bowl Pessary	SSP055 - 095	Ila
Urethra Pessary	UP045 - 100	Ila
Urethrabowl Pessary	USP055 - 090	Ila

Basic UDI-DI: 426024582GYNRPA5

SRN: DE-MF-000005474

complies with the requirements of Regulation (EU) 2017/745 and other EU Guidelines that are applicable.

Indication of the risk class and the conformity assessment procedure:

Class Ila in accordance with Annex VIII of Regulation (EU) 2017/745
Conformity assessment procedure according to Annex IX Chapters I + III

Name and identification number of the Notified Body:

DNV MEDCERT GmbH
Identification number: 0482

Certificate with validity:

Certificate ID: 1340GB448241217
Valid until: 2028-07-02

This declaration of conformity is valid until the above certificate expires or if a change is made to the product.

Department	Date:	Released by:
CEO / PRRC – Prof. Dr.med. Dr.h.c.mult. B.Arabin	02.01.2025	<i>B. Arabin</i>
QM Representative – M. Prior	02.01.2025	<i>Monika Prior</i>